



湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

证书编号: [REDACTED] 号

Certificate No.

- 本中心是国家法定计量检定机构,也是经合格评定国家认可委员会(CNAS)认可的实验室。

This body is an Institute of legal verification, it is also a accreditation laboratory accredited by China National Accreditation Service for Conformity Assessment(CNAS).

国家法定计量检定机构授权证书号: (国)法计(2012)01028, (国)法计(2012)01040

Authorization certificate number of the national legal metrological verification institution:

中国合格评定国家

数据/结果

Data/Results of Calibration

一、示值测量

测量参数	相别	被检显示值	实际值	测量不确定度 (U, k=2)
		57.7000V	57.6906V	0.0008V
		100.000V	99.9906V	0.002V
		220.000V	219.954V	0.002V
		380.000V	379.916V	0.010V
		57.70000V	57.6907V	0.00008V
		100.0000V	99.9907V	0.002V

		219.983V	219.983V	0.002V
		380.000V	379.936V	0.010V
		219.983V	219.983V	0.00002V

		219.983V	219.983V	0.002V
		220.000V	219.962V	0.0002V
		380.000V	379.922V	0.010V
		0.10000A	0.09998A	0.00002A
交流电流		1.50000A	1.50003A	0.00002A

50Hz		5.0000A	4.9999A	0.0002A
		10.0000A	9.9992A	0.0003A

测量参数	相别	被检显示值	实际值	测量不确定度 (U, k=2)
		0.10000A	0.09998A	0.00002A
		1.50000A	1.50002A	0.00002A

50Hz	0.10000A	0.09999A	0.00002A
	1.50000A	1.50002A	0.00002A