

湖北省计量测试技术研究院

Hubei Institute of Metrology and Testing Technology

中国合格评定服务联合体
China United Calibration Service

地址 (Add): 湖北省武汉市东湖新技术开发区茅 邮编 (Post Code): 430223

CUC



● 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

● 本次校准的技术依据（名称、代号）

Reference documents for the Calibration (Name、Code)

参照：JJG 874—2007 温度指示控制仪检定规程

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称 Name of Equipment	型号/编号 Model/Serial No.	证书号/有效期 Certificate No./Due Date
温湿度场自动测试系统	CST4002/20130703	2022WD03530001/2023-07-23

● 校准环境条件

Environmental condition on the Calibration

温度：20.4℃	相对湿度：55 %	其它：—
Temperature	R.H.	Others
气压：—	地点：本院水果湖基地B220	原始记录编号：—
Pressure	Place	Record No.

本校准结论：何也等送检样品的本次校准有效。

It's Effect That Results of the sample received by VGM is successful Calibration.

本证书经评审合格并加盖部分复制本证书。

校准数据/结果

Data/Results of Calibration

显示温度(℃)	146
实测温度(℃)	144.8
扩展不确定度	$U=0.8\text{ }^{\circ}\text{C}$ $k=2$

以下空白

RE