



三 8

第 442 号

计量器具名称

Name of Instrument

SF6 纯度仪

型号 / 规格

Type/Specification

HTCD-H

出厂编号

Serial No.

18051631354

制造单位

Manufacturer

武汉特高压电力科技有限责任公司

(校准专用章)

批准

Authorized by

职务

Position

副所长

校准

Calibrated by

核 验

Inspected by

校准日期

Calibration Date

2022 年

05 月

25 日

Year

Month

Day

计量检定机构授权证书号: (国)法计(2012)01022

电子信箱: jsmiguest@jsmi.com.cn

Authorization Certificate No.(2012)01022

E-mail

地址: 南京市仙林大学城仙林大道 95 号(总部)

网站: WWW.jsmi.com.cn

邮编: 210023

Add.No.95,wenlanRoad,XianLin University City,Nanjing(Headquarter)

Website

Post Code

电话: (025) 86435803

传真: (025) 86435803

Tel

Fax

国家法定计量检定机构, 计量授权证书号: (国) 法计(2012) 01022 号。
This laboratory is a national legal metrological verification institute. Authorization certificate NO.(2012)01022.

本院通过了中国合格评定国家认可委员会的认可。实验室认可证书号: CNAS L0841

This laboratory has been assessed by China National Accreditation Scheme for Conformity

Assessment(CNAS).

Certificate of laboratory accreditation :No.CNAS L0841.

本实验室出具的数据均可溯源至国家计量基准。

All data issued by laboratory are traceable to national measurement standards.

本次校准的技术依据:


Reference documents for the calibration:

QTD CC06-2012 《气体分析仪校准方法》

Name	Model	Serial No.	Date
气体测速仪校准标准装置			

without the approval of the laboratory.
2、本证书的结果仅对所校仪器有效。
The results are only responsible for the instrument.

3、本证书未加盖校准专用章无效。
This certificate will be invalid if it is issued without official seal.



中国认可
国际互认
校准
CALIBRATION
CNAS L0841

证书编号:
Certificate No.:
第 3 页 共 3 页
Page of

校准结果/说明 (续页)

Results of calibration and additional explanation(continued page)

1. 外观及工作正常性检查: 完好
2. 示值误差:

分组	标准值 (%)	测量值 (%)	示值误差 (%FS)
SF ₆ /N ₂	99.99	99.9	-0.1
SF ₆ /N ₂	99.99	99.9	-0.1

SF ₆ /N ₂	99.99	99.9	-0.1
SF ₆ /N ₂	86.0	86.0	0.0